

ALABAMA DEPARTMENT OF FORENSIC SCIENCES  
ADMINISTRATIVE CODECHAPTER 370-1-1  
CHEMICAL TEST FOR INTOXICATION

## TABLE OF CONTENTS

370-1-1-.01	Admissibility Rule For The Evidential Breath Alcohol Test Method
370-1-1-.02	Evidential Breath Alcohol Testing Instrument Inspections (Repealed 8/11/03)
370-1-1-.03	Blood, Urine And Other Bodily Substances
370-1-1-.04	Field Sobriety Screening Tests
370-1-1-.05	Field Sobriety Alcohol Screening Devices
370-1-1-.06	Field Sobriety Drug Screening Devices
370-1-1-A	Appendix A
370-1-1-.01	<u>Admissibility Rule For The Evidential Breath Alcohol Test Method.</u>

(1) Purpose. This rule addresses the statutory elements required for the admissibility of evidence obtained from Breath Alcohol Testing Instruments. This rule contains the METHOD, as referred to by §32-5A-194 Code of Ala. 1975, as amended, which insures and verifies that each individual Breath Alcohol Test is performed accurately and reliably. IN EVENT THAT ONE OF THESE REQUIREMENTS IS NOT FULFILLED, BREATH ALCOHOL TESTING EVIDENCE MAY STILL BE OFFERED THROUGH TRADITIONAL EVIDENTIARY PREDICATE as approved of in Ex parte Mayo, 652 So. 2d 201 (Ala. 1994).

(2) Definitions and Abbreviations.

(a) Draeger 7110 MKIII

1. *Acceptable Breath Sample.* For the purpose of this rule the subject must provide the Draeger Alcotest 7110 MKIII with a breath sample of sufficient volume and duration, as required by the software to be accepted for analysis. Software versions prior to and including version 3.24 require the subject to provide a minimum of 1.5 liters of breath, for a minimum of 4.5 seconds. Software versions 3.25 or later require the subject to provide a minimum of 1.3 liters of breath, for a minimum of 4.0 seconds.

2. *Accuracy Check Fail.* When the Calibration Check results are outside the programmed parameters of the Draeger Alcotest 7110 MKIII, the instrument will abort the testing sequence.

3. *Admin. Breath Temperature Correction.* Administrative Breath Temperature Correction can result in an administrative lowering of the reported Breath Alcohol Result. When the Draeger Alcotest 7110 MKIII determines a subject's Breath Temperature is not 34.0oC an adjustment to the reported value will be made. When the Breath Temperature is above 34.0oC the adjustment will result in a lower reported Breath Alcohol Concentration.

4. *Alcohol.* For the purpose of this rule, use of the term alcohol shall refer to ethanol or ethyl alcohol unless otherwise specified.

5. *Air Blank Check.* The Draeger Alcotest 7110 MKIII Breath Alcohol Testing Instrument verifies that the analytical pathway is free of contamination.

6. *Ambient Air Check.* The Draeger Alcotest 7110 MKIII Breath Alcohol Testing Instrument verifies that the air used to purge the analytical pathway was free of contamination.

7. *Ambient Air Fail.* The Draeger Alcotest 7110 MKIII will terminate a Breath Testing Sequence when the detector systems respond to ethanol or other substances in the room air.

8. *Blowing Not Allowed.* Only when the "PLEASE BLOW" message appears will the Draeger Alcotest 7110 MKIII accept a subject breath sample. Any breath sample provided to the Draeger Alcotest 7110 MKIII at any time other than when prompted to do so will not be accepted by the instrument.

9. *Blowing Time Too Short.* When a breath sample is provided to the Draeger Alcotest 7110 MKIII that is sufficient in volume but deficient in duration, as required by the software, the instrument will not accept the sample.

10. *Br-Temp Meas Not Ok.* Breath Temperature Measurement Not Ok occurs when the difference between the two Breath Temperature thermistor values is outside the specified range for a single breath sample.

11. *Breath Alcohol Test.* For the purpose of this rule, two (2) acceptable breath samples in conjunction with two diagnostic checks and two acceptable calibration checks shall constitute a Breath Alcohol Test.

12. *Calibration.* The purpose of the calibration is to define the appropriate parameters to allow the instrument to accurately quantitate alcohol in a breath sample. The

Draeger Alcotest 7110 MKIII does not calibrate itself at the time of the test; calibration is performed at the Alabama Department of Forensic Sciences laboratory and subsequently verified at the time of the test with dry gas standards.

13. *Calibration Check*. For the purpose of this rule, a calibration check is the automated process whereby the Draeger Alcotest 7110 MKIII samples and tests a known dry gas standard to verify the instrument's acceptable calibration.

14. *Certificate of Analysis*. For the purpose of this rule, the Certificate of Analysis is the document generated by the Draeger Alcotest 7110 MKIII for introduction of Breath Alcohol Test results to the Courts. By design, when the Certificate of Analysis indicates a numerical value for the subject's Breath Alcohol Concentration, the Certificate of Analysis is confirmation of successful method completion.

15. *Chemical Analysis*. For the purpose of this rule, chemical analysis is the separation of a substance into its constituent elements to determine either their nature or their portions. The goal as set forth in §32-5A-194 Code of Ala. 1975, as amended, is the chemical analyses of a person's breath to determine the presence and quantity of alcohol (ethanol).

16. *Data Pack*. The Data Pack contains the supporting data for the Certificate of Analysis as referred to at the bottom of that document. By design, the Data Pack along with the Subject Mis-Try file, and when needed the IR Curves, EC Profile Plot, and Flow Profile Plot, constitute "...full information concerning the test or tests shall be made available to him or his attorney..." as referred to by §32-5A-194(a)(4) Code of Ala. 1975, as amended.

17. *Draeger*. For the purpose of this rule, Draeger is the appropriate English translation of the German name Dräger.

18. *Deprivation Period*. For the purpose of the METHOD as referred to by §32-5A-194 Code of Ala. 1975, as amended, prior to submitting to a Breath Alcohol Test a person should not be allowed to put anything in their mouth for at least 20 minutes.

19. *EC*. For the purpose of this rule, EC will be the abbreviation for the electrochemical cell detector utilized by the Draeger Alcotest 7110 MKIII.

20. *EC Profile Plot*. For the purpose of this rule, the EC Profile Plot is a plot utilized by the Draeger Alcotest 7110 MKIII that depicts EC Detector signal for a given breath sample.

21. *Flow Profile Plot*. For the purpose of this rule, the Flow Profile Plot is a plot utilized by the Draeger Alcotest 7110 MKIII that depicts breath sample flow into the instrument.

22. *Forensic Edits*. In the event an operator incorrectly input clerical data as requested by the Breath Alcohol Testing method, this information can be corrected by the following procedure: the operator should strike through the incorrect data (preferably a single strike leaving the original data legible), insert the correct data, and initial and date the correction.

23. *Interference*. When the detector systems of the Draeger Alcotest 7110 MKIII respond to a substance other than alcohol (ethanol, ethyl alcohol), the instrument will terminate the test.

24. *IR*. For the purpose of this rule, IR will be the abbreviation for the infrared detector utilized by the Draeger Alcotest 7110 MK III.

25. *IR Curve*. For the purpose of this rule, the IR Curve is a plot of the infrared detector data utilized by the Draeger Alcotest 7110 MK III for a given breath sample.

26. *Method*. For the purpose of this rule a method is an orderly and systematic approach to accomplishing a goal. The goal as set forth in §32-5A-194 Code of Ala. 1975, as amended, is the chemical analyses of a person's breath to determine the presence and quantity of alcohol (ethanol). Therefore the method refers only to those tasks, manual or automated, which occur at the time of the breath test and result in accurately identifying and quantifying the amount of alcohol on a particular person's breath. The method in its entirety is performed and confirmed at the time of the breath test.

27. *Minimum Volume Not Achieved*. When a breath sample is provided to the Draeger Alcotest 7110 MKIII that is deficient in volume, as required by the software, the instrument will reject the sample.

28. *Mouth Alcohol*. When a breath sample contains a measurable amount of alcohol originating from the mouth, and is higher in concentration than the end expiratory air, the Draeger Alcotest 7110 MKIII will determine the sample to contain residual mouth alcohol. Once the

presence of residual mouth alcohol has been determined the instrument will suspend the testing sequence for twenty (20) minutes. The testing sequence will resume at the end of the twenty-minute wait preserving all clerical data entries.

29. *Out of Measuring Range.* When the alcohol concentration of a breath sample exceeds 0.45 g/210L the instrument will abort the testing sequence.

30. *Operator.* For the purpose of this rule, an operator is an employee of a law enforcement agency possessing a valid permit (active permit number) and who is in control of the Draeger Alcotest 7110 MKIII during a testing sequence. The operator also acts as a "FACT" witness with regard to instrument operation in any litigation arising from the breath test. As stated by the Court in *Stubstad v. City of Orange Beach*, 575 So.2d 1240 (Ala.Cr.App. 1991), the operator need not be an expert on the mechanical functioning of the instrument.

31. *Permit.* For the purpose of this rule, a permit is an electronic mechanism for controlling access to a Breath Alcohol Testing Instrument through a unique operator permit number. Confirmation of an operator's active status as well as pending expiration date is designated on the Certificate of Analysis.

32. *Purging.* The Draeger Alcotest 7110 MKIII Breath Alcohol Testing Instrument cycles fresh air through the instrument analysis path to cleanse the pathway of the last sample analyzed.

33. *Refusal.* The operator may elect to end an incomplete test by declaring the subject has refused to provide two complete and acceptable samples. Refusal indicates the operator believes the subject to be intentionally obstructing the collection of evidence via the breath sample. The Draeger Alcotest 7110 MKIII Breath Alcohol Testing Instrument cannot make the decision for the operator that any unacceptable subject breath sample or sequence of unacceptable subject breath samples constitutes a subject refusal. After each unacceptable subject breath sample the operator has the opportunity to determine by which of the following three options the testing sequence will proceed: <1> REFUSAL <2> UNABLE <3> RESTART TEST. In event circumstances require abruptly ending the testing sequence with a refusal, the operator may select <R> at the "PLEASE BLOW" prompt.

34. *Standard Gas Supply.* Failure of one of the two dry gas samples to be delivered to the Draeger Alcotest 7110

MKIII for analysis will result in the termination of the testing sequence.

35. *Subject "Mis-Try"*. A subject "Mis-Try" is a breath sample that failed to meet the acceptable criteria as determined by the Draeger Alcotest 7110 MKIII. A record of the attempted breath sample is preserved in the Subject "Mis-Try" file.

36. *Unable*. The operator may elect to end an incomplete test by declaring the subject is unable to provide two complete and acceptable samples. Unable indicates the operator believes the subject to be incapable of providing the required breath sample and thereby is not intentionally obstructing the collection of evidence via the breath sample. The Draeger Alcotest 7110 MKIII Breath Alcohol Testing Instrument cannot assess the subject's physical condition for the operator. After each unacceptable subject breath sample the operator has the opportunity to determine by which of the following three options the testing sequence will proceed: <1> REFUSAL <2> UNABLE <3> RESTART TEST.

(b) Intoximeters Intox DMT Dual Sensor

1. *Acceptable Breath Sample*. For the purpose of this rule the subject must provide the Intoximeters Intox DMT Dual Sensor with a breath sample of sufficient volume and duration, as required by the software to be accepted for analysis. The software on the Intox DMT requires the subject to provide a minimum of 1.3 liters of breath, for a minimum of 4.0 seconds.

2. *Admin. Breath Temperature Correction*. The Intoximeters Intox DMT Dual Sensor applies a correction to the results of every breath test to account for variations in breath temperature. This correction results in a lower reported result.

3. *Administrative Test/Admin Test*. Non evidential breath test to be used in instances other than incident to a DUI arrest. For example, a pre-release test performed on an individual arrested for DUI.

4. *Alcohol*. For the purpose of this rule, use of the term alcohol shall refer to ethanol or ethyl alcohol unless otherwise specified.

5. *Air Blank Check*. The Intoximeters Intox DMT Dual Sensor Breath Alcohol Testing Instrument verifies that the analytical pathway is free of contamination.

6. *Ambient Air Check*. The Intoximeters Intox DMT Dual Sensor Breath Alcohol Testing Instrument verifies that the air used to purge the analytical pathway is free of contamination.

7. *Ambient Fail*. The Intoximeters Intox DMT Dual Sensor Breath Alcohol Testing Instrument will abort the test if it detects ethanol or other contaminants in the room air during the initial purge.

8. *Blank Error*. The Intoximeters Intox DMT Dual Sensor Breath Alcohol Testing Instrument will abort the test if it detects ethanol or other contaminants in the analytical pathway.

9. *Breath Alcohol Test*. For the purpose of this rule, two (2) acceptable breath samples in conjunction with three diagnostic checks and two acceptable standard checks shall constitute a Breath Alcohol Test.

10. *Breath Result > .45 g/210L*. The Intoximeters Intox DMT Dual Sensor will abort the test if the result from a subject breath sample exceeds 0.450 g/210L.

11. *Calibration*. The purpose of calibration is to define the appropriate parameters to allow the instrument to accurately quantitate alcohol in a breath sample. The Intoximeters Intox DMT Dual Sensor does not calibrate itself at the time of the test; calibration is performed at the Alabama Department of Forensic Sciences laboratory and subsequently verified at the time of the test.

12. *Certificate of Analysis*. For the purpose of this rule, the Certificate of Analysis is the document generated by the Intoximeters Intox DMT Dual Sensor for introduction of Breath Alcohol Test results to the Courts. By design, when the Certificate of Analysis indicates a numerical value for the subject's Breath Alcohol Concentration, the Certificate of Analysis is confirmation of successful method completion.

13. *Chemical Analysis*. For the purpose of this rule, chemical analysis is the separation of a substance into its constituent elements to determine either their nature or their portions. The goal as set forth in §32-5A-194, Code of Ala. 1975, as amended, is the chemical analyses of a person's breath to determine the presence and quantity of alcohol (ethanol).

14. *Data Pack*. The Data Pack contains the supporting data for the Certificate of Analysis as referred to at the bottom of that document. By design the Data Pack along with the IR Curves, Flow Profile Plot, and the EC Profile

Plot constitute "...full information concerning the test or tests shall be made available to him or his attorney..." as referred to by §32-5A-194(a)(4), Code of Ala. 1975, as amended.

15. *Diagnostic Check*. The Intoximeters Intox DMT Dual Sensor Breath Alcohol Testing Instrument verifies that the operational parameters are within range with each test.

16. *Diagnostic Check Failed*. The Intoximeters Intox DMT Dual Sensor Breath Alcohol Testing Instrument will abort the test if one of the operational parameters falls outside the acceptable range.

17. *Deprivation Period*. For the purpose of the METHOD as referred to by §32-5A-194, Code of Ala. 1975, as amended, prior to submitting to a Breath Alcohol Test a person should not be allowed to put anything in their mouth for at least 20 minutes.

18. *EC*. For the purpose of this rule, EC will be the abbreviation for the electrochemical cell detector utilized by the Intoximeters Intox DMT Dual Sensor.

19. *EC Profile Plot*. For the purpose of this rule, the EC Profile Plot is a plot generated by the Intoximeters Intox DMT Dual sensor that depicts the EC detector signal for a given breath sample.

20. *Flow Profile Plot*. For the purpose of this rule, the Flow Profile Plot is a plot generated by the Intoximeters Intox DMT Dual Sensor that depicts breath sample flow into the instrument. This plot is included on the same graph as the IR Curves.

21. *Fuel Cell*. For the purpose of this rule Fuel Cell is another term used to describe the electrochemical cell detector utilized by the Intoximeters Intox DMT Dual sensor.

22. *Forensic Edits*. In the event an operator incorrectly input clerical data as requested by the Breath Alcohol Testing method, this information can be corrected by the following procedure: the operator should strike through the incorrect data (preferably a single strike leaving the original data legible), insert the correct data, and initial and date the correction.

23. *Interference Detected*. When the detector systems of the Intoximeters Intox DMT Dual Sensor respond to a substance other than alcohol (ethanol, ethyl alcohol), the instrument will abort the test.



24. *Invalid Sample.* When a breath sample contains a measurable amount of alcohol originating from the mouth, and is higher in concentration than the end expiratory air, the Intoximeters Intox DMT Dual Sensor will determine the sample to contain residual mouth alcohol and thus declare it invalid. Once the presence of residual mouth alcohol has been determined the instrument will suspend the testing sequence for twenty (20) minutes. The testing sequence will resume at the end of the twenty-minute wait preserving all clerical data entries.

25. *IR.* For the purpose of this rule, IR will be the abbreviation for the infrared detector utilized by the Intoximeters Intox Dual Sensor.

26. *IR Curve.* For the purpose of this rule, the IR Curve is a plot of the infrared detector signal generated by the Intoximeters Intox DMT Dual Sensor for a given breath sample. The plot is included on the same graph as the Flow Profile Plot.

27. *Method.* For the purpose of this rule a method is an orderly and systematic approach to accomplishing a goal. The goal as set forth in §32-5A-194, Code of Ala. 1975, as amended, is the chemical analyses of a person's breath to determine the presence and quantity of alcohol (ethanol). Therefore the method refers only to those tasks, manual or automated, which occur at the time of the breath test and result in accurately identifying and quantifying the amount of alcohol on a particular person's breath. The method in its entirety is performed and confirmed at the time of the breath test.

28. *Operator.* For the purpose of this rule, an operator is an employee of a law enforcement agency possessing a valid permit (active permit number) and who is in control of the Intoximeters Intox DMT Dual Sensor during a testing sequence. The operator also acts as a "FACT" witness with regard to instrument operation in any litigation arising from the breath test. As stated by the Court in *Stubstad v. City of Orange Beach*, 575 So.2d 1240 (Ala.Cr.App. 1991), the operator need not be an expert on the mechanical functioning of the instrument.

29. *Permit.* For the purpose of this rule, a permit is an electronic mechanism for controlling access to a Breath Alcohol Testing Instrument through a unique operator permit number. Confirmation of an operator's active status as well as pending expiration date is designated on the Certificate of Analysis.

30. *Purge Error.* When the Intoximeters Intox DMT Dual Sensor detects contaminants in the air used to purge the analytical pathway the test will be aborted.

31. *Purging.* The Intoximeters Intox DMT Dual Sensor Breath Alcohol Testing Instrument cycles fresh air through the instrument analysis path to cleanse the pathway of the last sample analyzed.

32. *Sample Agreement Error.* The Intoximeters Intox DMT Dual Sensor requires that the results from the two required samples agree within 0.020 g/210L. In the event that the results from two samples do not agree within 0.020 g/210L, the Intoximeters Intox DMT Dual Sensor will finish the test and start a new test retaining all clerical data.

33. *Standard Check.* The Intoximeters Intox DMT Dual Sensor verifies calibration at two different concentrations during the breath alcohol test.

34. *Standard Out of Range.* When the calibration verification results are outside the programmed parameters of the Intoximeters Intox DMT Dual Sensor, the instrument will abort the testing sequence.

35. *Refusal.* The operator may elect to end an incomplete test by declaring the subject has refused to provide two complete and acceptable samples. Refusal indicates the operator believes the subject to be intentionally obstructing the collection of evidence via the breath sample. The Intoximeters Intox DMT Dual Sensor Breath Alcohol Testing Instrument cannot make the decision for the operator that any unacceptable subject breath sample or sequence of unacceptable subject breath samples constitutes a subject refusal. The Intoximeters Intox DMT Dual Sensor Breath Alcohol Testing Instrument allows the subject three minutes to provide an adequate sample. If the subject fails to provide an adequate sample in the allotted time the Operator is prompted to select either <Refused> or <Unable>. In the event that circumstances require abruptly ending the testing sequence with a refusal, the operator may select <R> at the "PLEASE BLOW" prompt.

36. *Subject Mis-Try.* A subject "Mis-Try" is a breath sample that failed to meet the acceptable criteria as determined by the Intoximeters Intox DMT Dual Sensor. A record of the attempted breath sample is preserved on the Flow Profile Plot which is on the same graph as the IR Curves.

37. *Unable*. The operator may elect to end an incomplete test by declaring the subject is unable to provide two complete and acceptable samples. Unable indicates the operator believes the subject to be incapable of providing the required breath sample and thereby is not intentionally obstructing the collection of evidence via the breath sample. The Intoximeters Intox DMT Dual Sensor Breath Alcohol Testing Instrument cannot assess the subject's physical condition for the operator. The Intoximeters Intox DMT Dual Sensor Breath Alcohol Testing Instrument allows the subject three minutes to provide an adequate sample. If the subject fails to provide an adequate sample in the allotted time the Operator is prompted to select either <Refused> or <Unable>. In the event that circumstances require abruptly ending the testing sequence with a refusal, the operator may select <R> at the "PLEASE BLOW" prompt.

(3) Approved Evidential Breath Alcohol Instrument List.

(a) Alcotest 7110 MKIII, Draeger Safety, Inc., Durango CO. For the purpose of this rule, variations or enhancements that do not have any bearing on the alcohol measuring capability of the instrument, such as the addition of a modem designated as an Alcotest 7110 MKIII C, are approved.

(b) Intoximeters Intox DMT Dual Sensor, Intoximeters Inc., St. Louis, MO. For the purpose of this rule, variations or enhancements that do not have any bearing on the alcohol measuring capability of the instrument are approved.

(4) Approved Evidential Breath Alcohol Test Method.

(a) Draeger Alcotest 7110 MKIII

1. The method of operation of the Draeger Alcotest 7110 MKIII is software driven and controlled so as not to be influenced by operator technique. The method requires the operator to input the following clerical data (steps 1-18) as prompted. It is appropriate for the operator to enter a "-" for steps 6.(i), 6.(ii), 6.(iii), 6.(iv), 6.(v), and 10 - 16 when the requested information cannot be obtained. Omission of an entry in any field will prevent the completion of the Breath Alcohol Testing method.

(i) PERMIT NO. (NUMBER): The operator must enter and confirm his/her unique Draeger permit number.

(ii) <1> DUI <2> ADMIN <3> DEMO: The operator must select the type of testing routine to be administered to the subject.

(iii) ENTER DRY GAS STND (STANDARD) #1 PRESSURE: The operator must observe and record the pressure from the gas regulator gauge on dry gas standard #1 (0.020 g/210L ethanol standard). It should be noted that the gas pressure has no bearing on the analytical result obtained from the DUI testing routine.

(iv) ENTER DRY GAS STND (STANDARD) #2 PRESSURE: The operator must observe and record the pressure from the gas regulator gauge on dry gas standard #2 (0.080 g/210L ethanol standard). It should be noted that the gas pressure has no bearing on the analytical result obtained from the DUI testing routine.

(v) 20 MIN (MINUTE) DEPRIVATION PERIOD: The operator must confirm that the subject has been under the control of the arresting officer, the operator, and/or other employee of a law enforcement agency for a minimum of 20 minutes. Enter "Y" for yes or "N" for no.

(vi) ARREST OFFICER SAME AS OPERATOR: The operator must designate whether the breath test operator is or is not the arresting officer. Enter "Y" for yes or "N" for no. If the arresting officer is not the breath test operator then the following information identifying the arresting officer will also be required.

(I) ARREST OFFICERS LAST NAME: The operator must enter the last name of the arresting officer.

(II) ARREST OFFICERS FIRST NAME: The operator must enter the first name of the arresting officer.

(III) ARREST OFFICERS MIDDLE INITIAL: The operator must enter the middle initial of the arresting officer.

(IV) ARREST OFFICERS IDENTIFICATION NUMBER: The operator must enter the agency identification number of the arresting officer.

(V) ARREST OFFICERS AGENCY: The operator must enter the agency of the arresting officer.

(vii) TIME OF OFFENSE: The operator must enter the date and time of the offense using the following format <MM/DD/YYYY\_HH:MM>.

(viii) COUNTY OF OFFENSE: The operator must enter the county in which the offense occurred.

(ix) SUBJECT'S UTC (UNIFORM TRAFFIC COMPLAINT NUMBER): The operator must enter the subject's Uniform Traffic Complaint number.

(x) SUBJECT LAST NAME: The operator must enter the subject's last name.

(xi) SUBJECT FIRST NAME: The operator must enter the subject's first name.

(xii) SUBJECT MIDDLE INIT (INITIAL): The operator must enter the subject's middle initial.

(xiii) SUBJECT STREET ADDR/APT (ADDRESS/APARTMENT): The operator must enter the subject's street or apartment address.

(xiv) SUBJECT TOWN/CITY: The operator must enter the subject's town or city of residence.

(xv) SUBJECT STATE: The operator must enter the subject's state of residence using appropriate two-letter designation.

(xvi) SUBJECT DL# (DRIVER'S LICENSE NUMBER) OR SS# (SOCIAL SECURITY NUMBER): The operator must enter the subject's driver's license number, Alabama file number, social security number or Alabama I.D. number.

(xvii) SUBJECT <M/F> (MALE/FEMALE): The operator must designate the subject's gender using "M" for male and "F" for female.

(xviii) SUBJECT DOB (DATE OF BIRTH): The operator must designate the subject's date of birth using the format <MM/DD/YYYY>. An entry of 01/01/1900 will be used when the subject's date of birth cannot be obtained.

2. Upon completion of clerical data entry the Draeger Alcotest 7110 MKIII continues the Breath Alcohol Testing method by performing or displaying the following automated steps or messages.

(i) PURGING

(ii) AMBIENT AIR CHECK

(iii) AIR BLANK CHECK

(iv) ACCURACY CHECK (0.020 g/210L)

- (v) PURGING
- (vi) AMBIENT AIR CHECK
- (vii) AIR BLANK CHECK
- (viii) INSERT MOUTHPIECE, PRESS BUTTON (Operator Task)
- (ix) PLEASE WAIT
- (x) PLEASE BLOW (First Subject Sample)
- (xi) STOP
- (xii) REMOVE MOUTHPIECE
- (xiii) PURGING
- (xiv) AMBIENT AIR CHECK
- (xv) AIR BLANK CHECK
- (xvi) PLEASE WAIT
- (xvii) PURGING
- (xviii) AMBIENT AIR CHECK
- (xix) AIR BLANK CHECK
- (xx) INSERT MOUTHPIECE, PRESS BUTTON (Operator Task)
- (xxi) PLEASE WAIT
- (xxii) PLEASE BLOW (Second Subject Sample)
- (xxiii) STOP
- (xxiv) REMOVE MOUTHPIECE
- (xxv) PURGING
- (xxvi) AMBIENT AIR CHECK
- (xxvii) AIR BLANK CHECK
- (xxviii) ACCURACY CHECK (0.080 g/210L)
- (xxix) PURGING
- (xxx) AMBIENT AIR CHECK

(xxxi) AIR BLANK CHECK

(xxxii) RESULT

(xxxiii) DATA STORED

(b) Intoximeters Intox DMT Dual Sensor

1. The method of operation of the Intoximeters Intox DMT Dual Sensor is software driven and controlled so as not to be influenced by operator technique. The method requires the operator to input certain clerical data in the appropriate box shown on the instrument screen. (1-16 below) It is appropriate for the operator to enter a "-" when the requested information cannot be obtained. Omission of an entry in any field will prevent the completion of the Breath Alcohol Testing method.

(i) Permit Number or Password: The Operator must enter and confirm his/her unique permit number after pressing the Start Test button to proceed with a test. Alternatively one of the ADFS scientists can enter their permit number and a password which will allow them access to several functions including running an evidential breath alcohol test.

(ii) Scan Driver's License? The Intoximeters Intox DMT Dual Sensor comes equipped with a card reader. The operator can choose to scan the subject's driver's license at this time and the instrument will populate the corresponding fields automatically. In the event that the subject does not have a license the operator must enter the pertinent information in sections 11.-16.

(iii) Same as Operator: If the Arresting Officer is the same as the Operator then the Operator can check the box "Same as Operator" and the Operator's information will be automatically populated.

(iv) If the Arresting Officer is not the Operator then the Operator should input the following information.

(I) Name (First, Middle Initial, Last): Operator will enter the Arresting Officer's first name, middle initial, and last name here.

(II) Badge #: The Operator should input the Arresting Officer's badge number.

(III) Agency: The Operator should input the Arresting Officer's Agency.

(v) Offense Information: Date (MM/dd/yyyy); the Operator will input the date of offense here.

(vi) Offense Information: Time (24 Hour); the Operator will input the time of offense here using a 24-hour clock in the HH:mm format.

(vii) Offense Information: UTC; the Operator will input the ticket number here if it is known.

(viii) Offense Information: County; the Operator will select the county of offense from the drop-down menu.

(ix) Offense Information: 20 Min. Deprivation Period; the Operator must affirm that there has been 20 minutes between the time of offense and the time of test from the drop-down menu.

(x) Subject Information: Name (F/M/L); the Subject's first name, middle initial, and last name are entered here.

(xi) Subject Information: Street/Apt/Box; the subject's street address will be entered here.

(xii) Subject Information: City/Town; the city / town in which the subject resides is entered here.

(xiii) Subject Information: State; the state in which the subject resides is entered here.

(xiv) Subject Information: Sex; the Operator must select the gender of the subject from the pull-down menu.

(xv) Subject Information: Date of Birth (MM/dd/yyyy): The Subject's date of birth is entered here. An entry of 01/01/1900 will be used when the subject's date of birth cannot be obtained.

(xvi) Subject Information: License or SS Number; the Subject's Driver's License number or Social Security Number is entered here.

2. Upon completion of clerical data entry the Intoximeters Intox DMT Dual Sensor continues the Breath Alcohol Testing method by performing or displaying the following automated steps or messages.

(i) TIMED PURGE

(ii) AMBIENT ZEROING



- (iii) DIAGNOSTIC CHECK
- (iv) BLANK TEST
- (v) EXTERNAL STANDARD (0.020 g/210L)
- (vi) PURGING
- (vii) AMBIENT ZEROING
- (viii) BLANK TEST
- (ix) DID SUBJECT REFUSE? Y/N
- (x) INSERT MOUTHPIECE
- (xi) PLEASE BLOW
- (xii) REMOVE MOUTHPIECE
- (xiii) PURGING
- (xiv) DIAGNOSTIC CHECK
- (xv) AMBIENT ZEROING
- (xvi) BLANK TEST
- (xvii) DID SUBJECT REFUSE? Y/N
- (xviii) INSERT MOUTHPIECE
- (xix) PLEASE BLOW
- (xx) REMOVE MOUTHPIECE
- (xxi) PURGING
- (xxii) AMBIENT ZEROING
- (xxiii) BLANK TEST
- (xxiv) EXTERNAL STANDARD (0.080 g/210L OR 0.150 g/210L)
- (xxv) PURGING
- (xxvi) AMBIENT ZEROING
- (xxvii) BLANK TEST
- (xxviii) DIAGNOSTIC CHECK

(xxix) RESULT

(xxx) DATA STORED

(xxxi) PRINTOUT

(5) Report of Breath Alcohol Test Result.

(a) Draeger 7110 MKIII C

1. The Certificate of Analysis produced for each individual Breath Alcohol Test is confirmation of successful METHOD completion when an analytical result (number) is obtained. For a Breath Alcohol Test result to be reported, indicating successful METHOD completion, the following steps must be performed.

(i) "DIAGNOSTIC CHECKS BEFORE AND AFTER OK": An automated internal diagnostic check is performed by the Draeger Alcotest 7110 MKIII one hundred twenty-eight (128) times per second. For each Breath Alcohol Test, the Draeger Alcotest 7110 MKIII stores in memory a record of a single diagnostic "Snap Shot" before the first subject sample is collected and after the second subject sample is collected.

(ii) "CALIBRATION CHECKS BEFORE AND AFTER OK": An automated calibration check is performed before the first subject sample is collected and after the second subject sample is collected to verify the calibration of the Draeger Alcotest 7110 MKIII at the time of the breath test. A 0.020 g/210L ethanol standard is introduced into the instrument before the first subject sample is collected. An acceptable result of 0.015 to 0.025 g/210L will allow the Breath Alcohol Test to proceed. A 0.080 g/210L ethanol standard is introduced into the instrument after the second subject sample is collected. An acceptable result of 0.076 to 0.084 g/210L must be obtained before the results of the Breath Alcohol Test will be reported. When results obtained from the calibration checks are not within the stated acceptable ranges, the "Certificate of Analysis" obtained from the Draeger Alcotest 7110 MKIII will indicate ACCURACY CHECK FAIL and no numerical result will be reported at the time of the breath test.

(iii) "INSPECTION BEFORE AND AFTER TEST OK": Completed Diagnostic Checks and Calibration Checks verify instrument accuracy and reliability and therefore constitute a time of test inspection. When a condition exist which prevents the completion of either the diagnostic checks or the calibration

checks the "Certificate of Analysis" will not bear this statement.

(iv) "THE SUBJECT MUST COMPLETE A DEPRIVATION PERIOD OF AT LEAST TWENTY MINUTES BEFORE PROVIDING THE FIRST BREATH SAMPLE". An operator must attest that reasonable efforts were made by an employee of a law enforcement agency to deprive the subject of putting potentially interfering substances in their mouth. In event the operator becomes aware of potentially interfering substances in the mouth, the deprivation period should be restarted.

(v) "TWO ACCEPTABLE BREATH SAMPLES WERE ANALYZED, THE LOWEST RESULT IS REPORTED". The subject must provide two acceptable breath samples for analysis. The lowest alcohol result generated from each breath sample must agree within 0.020 g/210L of breath. If the alcohol results from the two breath samples do not agree within 0.020 g/210L of breath, the instrument will begin a second breath sample collection sequence. Failure to provide two (2) acceptable breath samples for the second breath sampling sequence will constitute a refusal of the whole test. When the results of two acceptable breath samples do agree within 0.020 g/210L of breath, the Draeger Alcotest 7110 MKIII will report the lowest result truncated to the second decimal place.

(vi) When the subject provides a breath sample that does not meet the minimum acceptable criteria or when not prompted to do so, one of the following errors will be recorded in the Subject "Mis-Try" file: "Minimum Volume Not Achieved", "Blowing Time Too Short", or "Blowing Not Allowed". When the Draeger Alcotest 7110 MKIII determines a sample or samples to be unacceptable the operator has the opportunity to determine by which of the following three options the testing sequence will proceed: <1> REFUSAL <2> UNABLE <3> RESTART TEST.

2. Four copies of the Certificate of Analysis will be generated each bearing the statement "SUPPORTING DATA FOR THIS DOCUMENT IS AVAILABLE UPON WRITTEN REQUEST TO THE ALABAMA DEPARTMENT OF FORENSIC SCIENCES, IMPLIED CONSENT UNIT". The supporting data or information as referred to in §32-5A-194(a)(4) Code of Ala. 1975, as amended, for the Certificate of Analysis consists of the "Data Pack", "Subject Mis-Try File", and when needed IR Curves, EC Profile Plot, and Flow Profile Plot.

(b) Intoximeters Intox DMT Dual Sensor

1. The Certificate of Analysis produced for each individual Breath Alcohol Test is confirmation of successful METHOD completion when an analytical result (number) is obtained. For a Breath Alcohol Test result to be reported, indicating successful METHOD completion, the following steps must be performed.

(i) "DIAGNOSTIC CHECKS BEFORE AND AFTER OK": An automated internal diagnostic check is performed by the Intoximeters Intox DMT Dual Sensor during every breath test. For each Breath Alcohol Test, the Intoximeters Intox DMT Dual Sensor stores in memory a record of a single diagnostic "Snap Shot" before the first subject sample is collected, between the first and second subject samples and after the second subject sample is collected.

(ii) "CALIBRATION CHECKS BEFORE AND AFTER OK": An automated calibration check is performed before the first subject sample is collected and after the second subject sample is collected to verify the calibration of the Intoximeters Intox DMT Dual Sensor at the time of the breath test. A 0.020 g/210L ethanol standard is introduced into the instrument before the first subject sample is collected. An acceptable result of 0.015 to 0.025 g/210L will allow the Breath Alcohol Test to proceed. A subject sample result of less than 0.150 g/210L will result in the introduction of a 0.080 g/210L ethanol standard into the instrument after the second subject sample is collected. An acceptable result of 0.076 to 0.084 g/210L must be obtained before the results of the Breath Alcohol Test will be reported. A subject sample result of 0.150 g/210L or higher will result in the introduction of a 0.150 g/210L ethanol standard into the instrument after the second subject sample is collected. An acceptable result of 0.1425 to 0.1575 g/210L must be obtained before the results of the Breath Alcohol Test will be reported. When results obtained from the calibration checks are not within the stated acceptable ranges, the "Certificate of Analysis" obtained from the Intoximeters Intox DMT Dual Sensor will indicate STANDARD OUT OF RANGE and no numerical result will be reported at the time of the breath test.

(iii) "INSPECTION BEFORE AND AFTER TEST OK": Completed Diagnostic Checks and Calibration Checks verify instrument accuracy and reliability and therefore constitute a time of test inspection. When a condition exist which prevents the completion of either the diagnostic checks or the calibration

checks the "Certificate of Analysis" will not bear this statement.

(iv) "THE SUBJECT MUST COMPLETE A DEPRIVATION PERIOD OF AT LEAST TWENTY MINUTES BEFORE PROVIDING THE FIRST BREATH SAMPLE". An operator must attest that reasonable efforts were made by an employee of a law enforcement agency to deprive the subject of putting potentially interfering substances in their mouth. In the event the operator becomes aware of potentially interfering substances in the mouth, the deprivation period should be restarted.

(v) "TWO ACCEPTABLE BREATH SAMPLES WERE ANALYZED, THE LOWEST RESULT IS REPORTED". The subject must provide two acceptable breath samples for analysis. The lowest alcohol result generated from each breath sample must agree within 0.020 g/210L of breath. If the alcohol results from the two breath samples do not agree within 0.020 g/210L of breath, the instrument will begin a second breath sample collection sequence. Failure to provide two (2) acceptable breath samples for the second breath sampling sequence will constitute a refusal of the whole test. When the results of two acceptable breath samples do agree within 0.020 g/210L of breath, the Intoximeters Intox DMT Dual Sensor will report the lowest result truncated to the second decimal place.

(vi) The Intoximeters Intox DMT Dual Sensor allows the subject 3 minutes to provide a sample. In the event of an inadequate sample the Operator may coach the subject to provide an acceptable sample. Any Mis-Try will be included on the Flow Profile Plot which is printed with the IR Curves at the time of the test.

2. Four copies of the Certificate of Analysis will be generated each bearing the statement "SUPPORTING DATA FOR THIS DOCUMENT IS AVAILABLE UPON WRITTEN REQUEST TO THE ALABAMA DEPARTMENT OF FORENSIC SCIENCES, IMPLIED CONSENT UNIT". The supporting data or information as referred to in §32-5A-194(a)(4) Code of Alabama, 1975 as amended, for the Certificate of Analysis consists of the "Data Pack", IR Curves, EC Profile Plot, and Flow Profile Plot (INCLUDES Mis-Tries). The IR Curves, EC Profile Plot, and Flow Profile Plot are printed at the time of the test.

(6) Qualifications: An applicant must satisfactorily complete a new operator's course in the operational procedures of the Breath Alcohol Testing Instrument and be an employee for one of the agencies listed in §32-5A-194, Code of Ala. 1975, as amended.

**(7) Permits:**

(a) Permits to perform a chemical analysis of a person's breath pursuant to §32-5A-194, Code of Ala. 1975, as amended, will be issued by the Director of the Department of Forensic Sciences upon the recommendation of the Technical Director of the Implied Consent Discipline of the Department of Forensic Sciences.

(b) Permits issued to new operators will automatically expire at the end of the succeeding calendar year.

(c) Operator permit status can remain active by satisfactorily completing a continuing education session each calendar year and by being an employee for one of the agencies listed in §32-5A-194, Code of Ala. 1975, as amended. The permit of an operator failing to complete a continuing education session any year after their new operator's course will automatically expire at the end of that calendar year.

(d) In instances where a declared State of Emergency was issued at the State or Federal level during the course of a calendar year, the continuing education requirement for that calendar year may be waived by the Director of the Department of Forensic Sciences or the Technical Director of the Implied Consent Discipline of the Department of Forensic Sciences. Successful operation and completion of a breath test resulting in a certificate analysis is confirmation of acceptable operator status, as referred to in Lum v. City of Brewton, 883 So.2nd 241 (Ala.Cr.App. 2003)

(e) An operator who fails to attend a continuing education session may be reactivated provided they attend a continuing education session within two calendar years. Any operator who remains inactive for more than two successive calendar years must attend a new operator's class to be reinstated.

(f) Any action or practice which is misleading or deceptive, or the violation of any of the rules of the Alabama Department of Forensic Sciences promulgated under the provisions of §32-5A-194, Code of Ala. 1975, as amended, shall constitute grounds upon which the Director may revoke such permit.

(g) If the Director receives a complaint or has reason to believe that an operator is participating in misleading or deceptive practices, violating or has violated any of the rules, he shall notify the operator of a hearing to determine if the alleged infraction has occurred. The Director will designate a place and time for the hearing.

(h) The Director or his designee shall conduct the hearing.

(i) Upon revocation of a permit, the Director shall notify the operator, the operator's immediate superior and the Technical Director of the Implied Consent Section of the Alabama Department of Forensic Sciences.

**Relevant Cites:**

**Designated Instrument:** *Harper v. City of Troy*, 467 So.2d 269 (Ala.Cr.App. 1985)

**Method:** *McDaniel v. State*, 706 So. 2d 1305, (Ala.Crim.App. 2001).

**Predicate:** *Ex parte Mayo*, 652 So. 2d 201 (Ala. 1994); *Ex Parte Vizzina*, 533 So.2d 269,271 (Ala. 1988); *Ex Parte Bush*, 474 So.2d 168 (Ala. 1985); *Moore v. State*, 442 So.2d 164, 167 (Ala.Cr.App. 1983); *Patton v. City of Decatur*, 337 So.2d 321 (Ala.1976); *McGough v. Slaughter*, 395 So.2d 972, 977 (Ala.1981).

**Testimony:** *Stubstad v. City of Orange Beach*, 575 So.2d 1240 (Ala.Cr.App. 1991)

**Author:** Gregory L. Turner, Mark A. Pevey, Dale A. Carpenter

**Statutory Authority:** Code of Ala. 1975, §§32-5A-194, as amended.

**History: New Rule:** Filed December 7, 1994; effective January 11, 1995. **Amended:** Filed October 10, 1995; effective November 11, 1995. (This function was transferred by Act No. 88-660 to the Department of Forensic Sciences, May 13, 1988.) **Amended:** Filed July 24, 1996; effective August 28, 1996. **Amended:** Filed December 4, 1998; effective January 8, 1999. **Amended:** Filed May 7, 1999; effective June 11, 1999. **Amended:** Filed July 7, 2003; effective August 11, 2003. **Amended:** Filed June 4, 2004; effective July 9, 2004. **Amended:** Filed May 10, 2018; effective June 24, 2018. **Amended:** Filed May 6, 2019; effective June 20, 2019. **Amended:** Published September 30, 2020; effective October 14, 2020.

**370-1-1-.02      Evidential Breath Alcohol Testing Instrument Inspections (Repealed 8/11/03).**

**Author:** Department of Forensic Sciences

**Statutory Authority:** Code of Ala. 1975, §§32-5A-194, as amended.

**History: New Rule:** Filed December 7, 1994; effective January 11, 1995. **Amended:** Filed October 10, 1995; effective November 11, 1995. (This function was transferred by Act No. 88-660 to the Alabama Department of Forensic Sciences, May 13, 1988.) **Amended:** Filed July 24, 1996; effective August 28, 1996. **Amended:** Filed December 4, 1998; effective January 8, 1999. **Amended:** Filed May 7, 1999; effective June 11, 1999. Repealed - Appendices A, B and D also: Filed July 7, 2003; effective August 11, 2003.

**370-1-1-.03      Blood, Urine And Other Bodily Substances.**

(1) Purpose. This rule describes the statutory elements required for the admissibility of evidence obtained from chemical analyses of a person's blood, urine, or other bodily substance. This rule contains the METHODS as referred to by §32-5A-194 Code of Ala. 1975, as amended, that insures and verifies each chemical analysis is performed accurately and reliably. IN THE EVENT THAT ONE OF THESE REQUIREMENTS IS NOT FULFILLED, CHEMICAL ANALYSIS EVIDENCE MAY STILL BE OFFERED THROUGH TRADITIONAL EVIDENTIARY PREDICATE as described in Powell v. State, 515 So.2d 140 (Ala. Cr. App. 1986).

(2) Definitions and Abbreviations:

(a) *Alcohol*- For the purpose of this rule, use of the term alcohol shall refer to ethanol or ethyl alcohol unless otherwise specified.

(b) *Method*- For the purpose of this rule, a method is an orderly and systematic approach to accomplishing a goal. The goal as set forth in §32-5A-194 Code of Ala. 1975, as amended, is the chemical analyses of a person's blood, urine or other bodily substance to determine the presence and/or quantity of alcohol (ethanol) or other chemical substances. Therefore the method refers only to those tasks, manual or automated, that result in accurately identifying and/or quantifying the amount of alcohol or other chemical substance in a particular person's blood or other bodily substance.

(c) *Permit*- For the purpose of this rule, a permit is authorization by the Director allowing an employee to analyze blood, urine, or other bodily substances for alcohol, drugs, or other chemical entities. By assigning a qualified employee this particular task the Director has expressed authorization.

(3) Qualifications. Applicants for a permit to perform a chemical analysis of a person's blood, urine, or other bodily substance pursuant to the Alabama Chemical Test for Intoxication Act shall meet the following requirements.

(a) Be employed as a Forensic Scientist by the Alabama Department of Forensic Sciences and be assigned to the Toxicology Section or Implied Consent Section by the Director.

(b) Have been approved by the Toxicology Discipline Chief to perform analyses on blood, urine, or other bodily substance analyses for the purpose of identification and quantitation of alcohol and/or other drugs.

(4) Certification Permits.



(a) Permits to perform a chemical analysis of a person's blood, urine, or other bodily substance pursuant to the Alabama Chemical Test for Intoxication Act will be issued by the Director and certified by the Toxicology Discipline Chief.

(b) Once issued, permits will remain effective concurrent with the term of employment of the individual. At the Directors discretion permits may be rescinded.

(5) Methods.

(a) Analysis of blood, urine, or other bodily substances for the presence of volatile compounds; such as alcohol (ethanol), shall be performed by Gas Chromatography utilizing internal standard headspace injection, in accordance with the Toxicology Discipline's Standard Operating Procedure for that analysis.

(b) Analysis of blood, urine, or other bodily substances for drugs or other chemical substances will utilize a screening technology in conjunction with Mass Spectrometry confirmation, in accordance with the Toxicology Discipline's Standard Operating Procedure for that analysis.

**Author:** Dale A. Carpenter, Jack R. Kalin, Curt E, Harper

**Statutory Authority:** Code of Ala. 1975, §32-5A-194, as amended.

**History: New Rule:** Filed December 7, 1994; effective January 11, 1995. **Amended:** Filed October 10, 1995; effective November 11, 1995. (This function was transferred by Act No. 88-660 to the Alabama Department of Forensic Sciences, May 13, 1988.) **Amended:** Filed July 24, 1996; effective August 28, 1996. **Amended:** Filed July 7, 2003; effective August 11, 2003. **Amended:** Filed May 10, 2018; effective June 24, 2018.

**370-1-1-.04      Field Sobriety Screening Tests.**

(1) Purpose. This rule describes the approved TECHNIQUE (Procedure) as referred to by §32-6-49.13 Code of Ala. 1975, as amended, that insures each Field Sobriety Screening Test performed on an operator of a commercial vehicle is performed in a standardized and reliable fashion.

(2) Definitions and Abbreviations. Alcohol

(a) - For the purpose of this rule, use of the term alcohol shall refer to ethanol or ethyl alcohol unless otherwise specified.

(3) Approved Procedure. The test battery comprising the Horizontal Gaze Nystagmus (HGN), One-leg Stand (OLS), and Walk and Turn (WAT) Field Sobriety Tests as described and applied according to the DUI Detection and Standardized Field Sobriety Testing Student Manual,

Publication HS178, Transportation Safety Institute, National Highway Traffic Safety Administration, U.S. Department of Transportation in its June 1992 or successor printings.

(4) Training Certificates. A certificate will be issued to each officer who successfully passes written and practical examinations during a minimum of 16 hours of Standardized Field Sobriety Test Training which included no less than two controlled drinking workshops using volunteer drinkers and wherein the administration and interpretation of the HGN, OLS, and WAT tests was presented.

**Author:** Dale A. Carpenter, Mark A. Pevey, Gregory L. Turner

**Statutory Authority:** §32-6-49.13, as amended.

**History: New Rule:** Filed July 24, 1996; effective August 28, 1996. **Amended:** Filed July 7, 2003; effective August 11, 2003.

### **370-1-1-.05      Field Sobriety Alcohol Screening Devices.**

(1) Purpose. This rule list the approved "FIELD BREATHALYZER or OTHER APPROVED DEVICE" as referred to by §32-6-49.13 Code of Ala. 1975, as amended. Additionally, this rule describes training requirements and minimal operational criteria necessary for accurate and reliable results.

(2) Definitions and Abbreviations:

(a) *Alcohol*. For the purpose of this rule, use of the term alcohol shall refer to ethanol or ethyl alcohol unless otherwise specified.

(b) *Observation*. For the purpose of this rule, use of the term observation shall mean to watch. Prior to the administration of a field sobriety screening test employing an approved field sobriety screening device a subject must be under the observation of the operator of the device or other law enforcement officer for a period of not less than twenty minutes.

(3) Approved Training. Training afforded by the manufacturer of an approved device and/or training received as a part of the Alabama Preliminary Breath Testing Instrument course shall constitute approved training.

(4) Operation of Devices.

(a) Officers shall use the device according to the manufacturer's operational procedure.

(b) Every subject must be under observation by the arresting officer for a period of twenty minutes before the screening device test is administered.

(5) Calibration. Calibration shall be checked every twelve (12) months. The device passes the calibration check if it renders a reading between 0.030 g/210L and 0.050 g/210L inclusive on a 0.040 g/210L percent standard delivered from either a wet bath simulator or a dry gas cylinder.

(6) Training and Calibration Records. It shall be the responsibility of each law enforcement agency to maintain permanent records documenting the training of each officer in the use of approved field sobriety screening devices and the annual calibration check results on each device in use by the law enforcement agency.

(7) Approved Field Sobriety Alcohol Screening Device List.

**NOTE:** For the purpose of this rule, variations or enhancements that do not have any bearing on the alcohol measuring capability of the instrument, such as the addition of a modem, external printer or passive sampling systems are approved.

The list of field sobriety alcohol screening devices approved for use in Alabama can be found by following the link on the Department's web page at [www.adfs.alabama.gov](http://www.adfs.alabama.gov).

**Author:** Dale A. Carpenter, Mark A. Pevey, Curt E. Harper, Gregory L. Turner

**Statutory Authority:** Code of Ala. 1975, §32-6-49.13, as amended.

**History: New Rule:** Filed July 24, 1996; effective August 28, 1996. **Amended:** Filed July 7, 2003; effective August 11, 2003.

**Amended:** Filed May 10, 2018; effective June 24, 2018.

### **370-1-1-.06      Field Sobriety Drug Screening Devices.**

(1) Purpose. This rule list the approved "FIELD BREATHALYZER or OTHER APPROVED DEVICE" as referred to by §32-6-49.13, Code of Ala. 1975, as amended. Additionally, this rule describes training requirements and minimal operational criteria necessary for accurate and reliable results from oral fluid drug screening devices.

(2) Definitions and Abbreviations:

(a) Drug. Any substance, when taken into the human body, which can impair the ability of a person to operate a vehicle safely.

(b) Negative Result. A negative result indicates the samples is drug-free for the tested targets or below the cutoff level of the test.

(c) Observation. For the purpose of this rule, use of the term observation shall mean to watch. Prior to the administration

of a field sobriety screening test employing an approved field sobriety drug screening device a subject must be under the observation of the operator of the device or other law enforcement officer for a period of not less than ten minutes.

(d) Oral Fluid. A clear, tasteless fluid comprised of saliva produced by multiple salivary glands, and other constituents inside the mouth.

(e) Positive Result. A positive result indicates presence of the drug, its metabolites, or cross-reacting substance but does not indicate level of intoxication, administration route or concentration in the oral fluid. A positive test result should be confirmed by a second test LC/MS (liquid chromatography-mass spectrometry).

(3) Approved Training. Training afforded by the manufacturer of an approved device and/or training received from the Alabama Department of Forensic Sciences shall constitute approved training.

(4) Operation of Devices.

(a) Officers shall use the device according to the manufacturer's operational procedure.

(b) Every subject must be under observation by an officer for a period of ten minutes before the screening device test is administered.

(c) The subject should not eat, drink, or smoke ten minutes prior to giving a sample.

(5) Quality Control Tests and Maintenance.

(a) Quality control (QC) tests and annual maintenance shall be conducted per manufacturer's operational procedure.

(b) The device is working properly if the QC test(s) pass.

(c) QC test(s) should be conducted at the time of testing or within 24 hours of the subject test.

(6) Training and Maintenance Records. It shall be the responsibility of each law enforcement agency to maintain permanent records documenting the training of each officer in the use of approved field sobriety screening devices and the manual maintenance results on each device in use by the law enforcement agency.

(7) Approved Field Sobriety Drug Screening Device List.

NOTE: For the purpose of this rule, variations or enhancements that do not have any bearing on the drug detecting capability of the instrument, such as the addition of a modem, external printer or passive sampling systems are approved.

The list of field sobriety alcohol screening devices approved for use in Alabama can be found by following the link on the Department's web page at [www.adfs.alabama.gov](http://www.adfs.alabama.gov).

**Author:** Curt E. Harper

**Statutory Authority:** Code of Ala. 1975, §32-6-49.13, as amended.

**History: New Rule:** Filed May 10, 2018; effective June 24, 2018.

## Appendix A

**EVIDENTIAL BREATH ALCOHOL TESTING QUALITY CONTROL.**

## Purpose.

The purpose of this section is to inform the public of the quality control or good laboratory practices that are utilized by the Alabama Department of Forensic Sciences to support Breath Alcohol Testing in the field. The practices described in this section only demonstrate that each Draeger Alcotest 7110 MKIII is capable of performing as expected. The method as described in 370-1-1-.01(4) is the only sequence of steps that isolates alcohol (ethanol) from a subject to accurately determine the amount. Subsequently the quality control built into the method is the only process that verifies a particular Breath Alcohol Test did indeed perform as expected. Since the authority derived from §32-5A-194 Code of Ala. 1975, as amended, is limited to a method to perform a chemical analysis to determine the alcoholic content of a subjects blood the following information pertaining to good laboratory practice **does not constitute a rule.**

(1) Pre-installation Evaluation. Each Draeger Alcotest 7110 MKIII is evaluated by the Alabama Department of Forensic Sciences Technical Director or his designee prior to being placed in operation. The evaluation will demonstrate the ability of each Draeger Alcotest 7110 MKIII to identify and flag specific conditions; as well as, verify the analytical integrity of the instrument.

(a) *Purging Error.* Place the instrument in the "Error Check" mode. Restrict the air flow into the ambient air inlet while the pump is running. This will prompt the instrument to display "Check Sampling Hose" and sound an audible alarm. Record pass if the message "Error Triggered, Test Okay" is printed. If the error is not triggered initiate the necessary corrective actions to restore the instrument to proper working condition.

(b) *Minimum Volume Not Achieved.* Place the instrument in the "Error Check" mode. At the prompt "Please Blow", deliver a deficient sample volume. Record pass if the message "Error Triggered, Test Okay" is printed. If the error is not triggered initiate the necessary corrective actions to restore the instrument to proper working condition.

(c) *Blowing Time Too Short.* Place the instrument in the "Error Check" mode. At the prompt

"Please Blow", deliver a sample of sufficient volume and deficient duration. Record pass if the message "Error Triggered, Test Okay" is printed. If the error is not triggered initiate the necessary corrective actions to restore the instrument to proper working condition.

(d) *Blowing Not Allowed.* Place the instrument in the "Error Check" mode. At any time other than when the instrument indicates "Please Blow" deliver a sample into the instrument. Record pass if the message "Error Triggered, Test Okay" is printed. If the error is not triggered initiate the necessary corrective actions to restore the instrument to proper working condition.

(e) *Ready To Blow Expired.* Place the instrument in the "Error Check" mode. At the prompt "Please Blow" do not provide a sample. After three (3) minutes has elapsed this error will be triggered by the instrument. Record pass if the message "Error Triggered, Test Okay" is printed. If the error is not triggered initiate the necessary corrective actions to restore the instrument to proper working condition.

(f) *Ambient Air Check.* Place the instrument in the "Error Check" mode. When the instrument begins to purge itself, direct an ethanol vapor (typically from a 0.020 g/210L dry gas cylinder) near the breath hose inlet, but not directly into the breath hose inlet. Record pass if the message "Error Triggered, Test Okay" is printed. If the error is not triggered initiate the necessary corrective actions to restore the instrument to proper working condition.

(g) *Interference.* Place the instrument in the "Error Check" mode. Prepare a methanol control by adding 105 microliters of methanol to 500 milliliters of distilled water and dispense into a simulator. At the prompt "Please Blow", deliver a sufficient sample into the instrument. Record pass if the message "Error Triggered, Test Okay" is printed. If the error is not triggered initiate the necessary corrective actions to restore the instrument to proper working condition.

(h) *Mouth Alcohol.* Place the instrument in the "Error Check" mode. At the prompt "Please Blow" the evaluator should rinse his or her mouth with a common mouthwash containing ethanol, then deliver a sufficient sample into the instrument. Record pass if the message "Error Triggered, Test Okay" is printed. If the error is not triggered initiate the necessary corrective actions to restore the instrument to proper working condition.

(i) *Accuracy Check Fail.* Place the instrument in the "Error Check" mode. Connect the 0.020 g/210L dry gas cylinder to the 0.080 g/210L gas port and initiate a test. Record pass if the message "Error Triggered, Test Okay" is printed. If the error is not triggered initiate the necessary corrective actions to restore the instrument to proper working condition.

(j) *Standard Gas Supply.* Disconnect the dry gas cylinders from the instrument and initiate a test. Record pass if the message "Std Gas Supply" is displayed. If the error is not triggered initiate the necessary corrective actions to restore the instrument to proper working condition.

(k) *Outside +/- Tolerance.* Initiate an administrative test sequence. At the prompt "Please Blow" deliver a sufficient sample. When prompted to deliver the second sample, deliver a sufficient sample which varies in concentration from the first by at least 0.021 g/ 210 L. Record pass if the message "Outside +/- Tol" is displayed. If the error is not triggered initiate the necessary corrective actions to restore the instrument to proper working condition.

(l) *No Admittance.* The evaluator should attempt to initiate a restricted function without the use of an appropriate key. Record pass if the message "No Admittance" is displayed. If the error is not triggered initiate the necessary corrective actions to restore the instrument to proper working condition.

(m) *Linearity Check.* Utilizing manufactured solutions prepare simulators with the following ethanol concentrations: 0.000g/210L, 0.020g/210L (+/- 0.005 g/210L), 0.040 g/210L (+/- 0.005 g/210L), 0.080 g/210L (+/- 5%), 0.120 g/210L (+/- 5%), 0.200 g/210L (+/- 5%), and 0.500 g/210L (+/- 5%). Place the instrument in the "Acc-Check" mode. When instructed, connect the appropriate simulator to the instrument. Repeat this procedure three times for each ethanol control. Average the results obtained for each concentration from the three runs. To record pass, the average results must fall within the following parameters: 0.000 g/210L, 0.020 g/210L (+/- 0.005 g/210L), 0.040 g/210L (+/- 0.005 g/210L), 0.080 g/210L (+/- 5%), 0.120 g/210L (+/- 5%), 0.200 g/210L (+/- 5%), and 0.500 g/210L (+/- 5%). If the instrument does not pass, initiate the necessary corrective actions to achieve the desired specifications.

(n) *Breath Temperature Check.* Place the instrument in "ABA" mode. Initiate a testing sequence consisting of three breath samples. When instructed to



do so deliver three breath samples into the instrument from a precisely controlled water bath apparatus adjusted to a temperature of approximately 34.0°C. Repeat this procedure using a precisely controlled water bath apparatus adjusted to a temperature of approximately 37.0°C. Record pass if the instrument records a temperature within +/- 0.3°C of the actual temperature. If the instrument does not pass initiate the necessary corrective actions to restore the instrument to proper working condition.

(o) *Acetone Interference Check.* Prepare a solution consisting of 500 milliliters of a manufactured 0.080 g/210L solution of ethanol and 950 microliters of Acetone in a simulator. Initiate an administrative test sequence. At the prompt "Please Blow", deliver a sufficient sample into the instrument. Record pass if the message "Interference" is displayed or if the instrument reports a result of 0.080 g/210 L +/- 0.004 g/210 L. If the check does not pass initiate the necessary corrective actions to restore the instrument to proper working condition.

(p) *Voltage Range Check.* Verify the Draeger Alcotest 7110 MKIII's performance capability throughout a voltage range of approximately 90-200 volts AC and 12 volts DC.

1. With the instrument connected to a variable AC power supply, adjust the voltage to approximately 90 volts AC. Initiate a standard check sequence consisting of three measurements of a 0.080 g/210L dry gas standard. Record pass if the instrument records an average result between 0.076-0.084 g/210L. If the instrument does not pass, initiate the necessary corrective actions to achieve the desired specification.

2. With the instrument connected to a variable AC power supply, adjust the voltage to approximately 220 volts AC. Initiate a standard check sequence consisting of three measurements of a 0.080 g/210L dry gas standard. Record pass if the instrument records an average result between 0.076-0.084 g/210L. If the instrument does not pass, initiate the necessary corrective actions to achieve the desired specification.

3. With the instrument connected to a DC power supply, adjust the voltage to approximately 12 volts DC. Initiate a standard check sequence consisting of three measurements of a 0.080 g/210L dry gas standard. Record pass if the instrument

records an average result between 0.076-0.084 g/210L. If the instrument does not pass initiate the necessary corrective actions to achieve the desired specification.

(q) *Barometer Check.* Establish communication with the instrument in the diagnostic screen mode. Compare the barometric pressure as indicated by the instrument to the laboratories barometer. Record pass if the difference between the two readings is not greater than 5%. If the instrument does not pass initiate the necessary corrective actions to restore the instrument to proper working condition.

(r) *Standard Deviation Check.* Initiate a standard check sequence consisting of ten measurements of a 0.080% dry gas standard. Record pass if the instrument records the following: an average result between 0.076-0.084 g/210L and a standard deviation of less than 0.0025. If the instrument does not pass initiate the necessary corrective actions to restore the instrument to proper working condition.

(2) *Data Download Review.* The Test Data collected and stored by each Draeger Alcotest 7110 MKIII in the field is transferred to the Alabama Department of Forensic Sciences. Upon transfer, the data is systematically reviewed to identify instrument problems and to perform trend analyses. This data transfer allows the Alabama Department of Forensic Sciences to review essentially 100% of the Breath Alcohol Tests performed statewide.

(a) *Communication Check.* The Draeger Alcotest 7110 Data Retrieval and Archiving Program produces a Summary of Automatic Data Retrieval. The summary readily identifies those instruments for which communication was not established. For those instruments that automatic communication was unsuccessful, manually establish communication. If communication problems persist investigate the source of the problem and document.

(b) *Message Check.* Open the Message File (50 File) to identify the automatic Instrument Messages and Operator Messages retrieved by the current download. Review the messages to determine the needs of the instrument, operator, or location (supplies). Document any actions taken.

(c) *Instrument Parameter Check.* Open the Data File (10 File) to identify Instrument Parameter data retrieved by the current download. Isolate the data corresponding to each of the following parameters collected before and after the subject samples for

review: *Ambient Air Pressure, 12V DC Power Supply, Battery Voltage, IR Signal, EC Offset, Cuvette Temperature, Breath Hose Temperature, Breath Probe Temperature, Breath Temperature Thermistors, Flow Sensor, Pre-Test Diagnostic Check, and Post-Test Diagnostic Check.* For each parameter identify when if any an instrument recorded a result outside the specified operating range. Investigate the source of any problem and document any corrective action.

(d) *Dry Gas Configuration Check.* Open the Data File (10 File) to identify the Dry Gas Configuration data retrieved by the current download. Isolate the data corresponding to each of the following categories for review: *0.02 Gas Lot Number, 0.02 Gas Expiration Date, 0.08 Gas Lot Number, 0.08 Gas Expiration Date, 0.02 Target Concentration, 0.02 Relative Tolerance, 0.02 Absolute Tolerance, 0.08 Target Concentration, 0.08 Relative Tolerance, and 0.08 Absolute Tolerance.* For each category identify if the instrument contains the correct configuration. Investigate any discrepancy and document any corrective action.

(e) *Error and Accuracy Check.* Open the Data File (10 File) to identify the Error and Accuracy Check data retrieved by the current download. Isolate the data corresponding to each of the following errors or measurements: *Error, Error 2, Error 3, Error 4, 0.02 gas data, and 0.08 gas data.* For each category identify when if any an instrument recorded a result outside the specified operating range. Also review the 0.02 g/210L and 0.08 g/210L gas data to identify low or high trends. Investigate the source of any problem or trend and document any corrective action.

(3) *Electronic Inspection.* The Electronic Inspection serves primarily as a quality assurance role by being a back-up inspection to the instrument set-up procedures and Data Download Review. The Electronic Inspection also is an opportunity to evaluate instrument performance over an extended time frame.

(a) *Diagnostic Screen Check.* Initiate the Draeger Alcotest 7110 Data Retrieval and Archiving Program. Establish communication with the instrument through the Diagnostic Screen. Evaluate the following displayed parameters: *Instrument Serial Number, Instrument Location, Firmware Version, Date, Time, Ambient Air Pressure, 12VDC Power Supply, Battery Voltage, IR Signal, EC Offset, Cuvette Temperature, Breath Hose Temperature, Breath Probe Temperature, Breath Temperature 1, Breath Temperature 2, Function Key, and*

*Flow Sensor.* Investigate any discrepancy and document any corrective action.

(b) *Interactive Screen Check.* Initiate the Draeger Alcotest 7110 Data Retrieval and Archiving Program. Establish communication with the instrument through the Interactive Screen. Evaluate the following displayed parameters: *Measuring Units, Calibration Configuration, and Cylinder Pressure Limit.* Investigate any discrepancy and document any corrective action.

(c) *Calibration Check Plot.* Open the Data File (10 File) to identify the Accuracy Check data retrieved by the instrument over the last 6 months. Plot the Accuracy Check data for both the 0.02 and 0.08 g/210L Accuracy Checks. Investigate the source of any problem or trend and document any corrective action.

(4) *Annual Re-Evaluation/Pre-Installation Evaluation.* Regardless of the performance of an instrument in the field each instrument will be brought back to the Alabama Department of Forensic Sciences laboratory once a year for evaluation. This evaluation will consist of the battery of tests described under (1) Pre-Installation Evaluation.

(a) *Intoximeters Intox DMT Dual Sensor*

1. *Pre-installation Evaluation.* Each Intoximeters Intox DMT Dual Sensor is evaluated by the Alabama Department of Forensic Sciences Technical Director or his designee prior to being placed in operation. The evaluation will demonstrate the ability of each Intoximeters Intox DMT Dual Sensor to identify and flag specific conditions; as well as, verify the analytical integrity of the instrument.

(i) *Purging Error.* While in the "Maintenance Test" mode, the instrument will ask "Trigger following test condition?" Purging Error. Restrict the air flow into the breath tube while the pump is running. This will prompt the instrument to sound an audible alarm. If the error is triggered, the instrument will record PASS and the message "Successfully induced a purging error" is displayed on the screen. If the error is not triggered initiate the necessary corrective actions to restore the instrument to proper working condition.

(ii) *Ready to Blow Expired.* While in the "Maintenance Test" mode, the instrument will ask "Trigger following test condition?" Ready to Blow Expired. At the prompt "Please Blow", never

provide an acceptable sample. If the error is triggered, the instrument will record PASS and the message "Successfully induced a ready to blow expired" is displayed on the screen. If the error is not triggered initiate the necessary corrective actions to restore the instrument to proper working condition.

(iii) *Ambient Fail.* While in the "Maintenance Test" mode, the instrument will ask "Trigger following test condition?" Ambient Fail. During the initial purge introduce a 0.020 g/210L ethanol dry gas into the breath hose. This process will be required 3 times. If the error is triggered the instrument will record PASS and the message "Successfully induced ambient fail" is displayed on the screen. If the error is not triggered initiate the necessary corrective actions to restore the instrument to proper working condition.

(iv) *Acetone Interference.* Prepare a solution consisting of 500 milliliters of a manufactured 0.080 g/210L solution of ethanol and 950 microliters of Acetone in a simulator. Allow the simulator to come to temperature. While in the "Maintenance Test" mode, the instrument will ask "Trigger following test condition?" Acetone Interference. When prompted, deliver a sufficient sample into the instrument through the simulator. If the error is triggered the instrument will record PASS and the message "Successfully induced acetone interference" is displayed on the screen. If the error is not triggered initiate the necessary corrective actions to restore the instrument to proper working condition.

(v) *Methanol Interference.* Prepare a methanol control by adding 105 microliters of methanol to 500 milliliters of distilled water and dispense into a simulator. Allow the simulator to come to temperature. While in the "Maintenance Test" mode, the instrument will ask "Trigger following test condition?" Methanol Interference. When prompted, deliver a sufficient sample into the instrument through the simulator. If the error is triggered the instrument will record PASS and the message "Successfully induced methanol interference" is displayed on the screen. If the error is not triggered initiate the necessary corrective

actions to restore the instrument to proper working condition.

(vi) *Invalid Sample.* While in the "Maintenance Test" mode, the instrument will ask "Trigger following test condition?" Invalid Sample. The evaluator should rinse his or her mouth with a common mouthwash containing ethanol. When prompted deliver a sufficient sample into the instrument. If the error is triggered the instrument will record PASS and the message "Successfully induced invalid sample" is displayed on the screen. If the error is not triggered initiate the necessary corrective actions to restore the instrument to proper working condition.

(vii) *Sample Agreement Error.* While in the "Maintenance Test" mode, the instrument will ask "Trigger following test condition?" Sample Agreement Error. When prompted, deliver a sufficient sample. When prompted to deliver the second sample, deliver a sufficient sample which varies in concentration from the first by at least 0.021 g/ 210 L. If the error is triggered the instrument will record PASS and the message "Successfully induced sample agreement error" is displayed on the screen. If the error is not triggered initiate the necessary corrective actions to restore the instrument to proper working condition.

(viii) *Standard Out of Range.* While in the "Maintenance Test" mode, the instrument will ask "Trigger following test condition?" Standard Out of Range. Connect an incorrect gas to a gas port and select that port, ca. connect a 0.080 g/ 210L gas to the 0.020 g/210L port. If the error is triggered the instrument will record PASS and the message "Successfully induced standard out of range" is displayed on the screen. If the error is not triggered initiate the necessary corrective actions to restore the instrument to proper working condition.

(ix) *Gas Flow Error.* While in the "Maintenance Test" mode, the instrument will ask "Trigger following test condition?" Gas Flow Error. When the gas standard box appears select a gas port that does not have a gas connected to it. If the error is triggered the instrument will record PASS and the message "Successfully induced gas flow error" is displayed on the

screen. If the error is not triggered initiate the necessary corrective actions to restore the instrument to proper working condition.

(x) *Linearity Check.* Utilizing manufactured solutions prepare simulators with the following ethanol concentrations: 0.000g/210L, 0.020g/210L (+/- 0.005 g/210L), 0.040 g/210L (+/- 0.005 g/210L), 0.080 g/210L (+/- 5%), 0.120 g/210L (+/- 5%), 0.200 g/210L (+/- 5%), and 0.500 g/210L (+/- 5%). Place the instrument in the "Linearity Test" mode. When instructed, connect the appropriate simulator to the instrument. Repeat this procedure three times for each ethanol control. Average the results obtained for each concentration from the three runs. To record pass, the average results must fall within the following parameters: 0.000 g/210L, 0.020 g/210L (+/- 0.005 g/210L), 0.040 g/210L (+/- 0.005 g/210L), 0.080 g/210L (+/- 5%), 0.120 g/210L (+/- 5%), 0.200 g/210L (+/- 5%), and 0.500 g/210L (+/- 5%). If the instrument does not pass, initiate the necessary corrective actions to achieve the desired specifications.

(xi) *Voltage Range Check.* The Intoximeters Intox DMT Dual Sensor utilizes and external power supply which supplies 12 volts DC to the instrument. Verify the Intoximeters Intox DMT Dual Sensor performance capability throughout a voltage range of approximately 90-200 volts AC.

(I) With the instrument connected to a variable AC power supply, adjust the voltage to approximately 90 volts AC. Initiate a standard check sequence consisting of three measurements of a 0.080 g/210L dry gas standard. Record pass if the instrument records an average result between 0.076-0.084 g/210L. If the instrument does not pass, initiate the necessary corrective actions to achieve the desired specification.

(II) With the instrument connected to a variable AC power supply, adjust the voltage to approximately 220 volts AC. Initiate a standard check sequence consisting of three measurements of a 0.080 g/210L dry gas standard. Record pass if the instrument records an average result between 0.076-0.084 g/210L. If the instrument does not pass, initiate the necessary corrective actions to achieve the desired specification.

(xii) *Barometer Check.* Place the instrument in "Technician Mode". Compare the barometric pressure as indicated by the instrument to the laboratory's barometer. Record pass if the difference between the two readings is not greater than 5%. If the instrument does not pass initiate the necessary corrective actions to restore the instrument to proper working condition.

(xiii) *Standard Deviation Check.* Initiate a standard check sequence consisting of ten measurements of a 0.080 g/210L dry gas standard. Record pass if the instrument records the following: an average result between 0.076-0.084 g/210L and a standard deviation of less than 0.0025. If the instrument does not pass initiate the necessary corrective actions to restore the instrument to proper working condition.

2. *Data Download Review.* The Test Data collected and stored by each Intoximeters Intox DMT Dual Sensor in the field is transferred to the Alabama Department of Forensic Sciences. Upon transfer, the data is systematically reviewed to identify instrument problems and to perform trend analyses. This data transfer allows the Alabama Department of Forensic Sciences to review essentially 100% of the Breath Alcohol Tests performed statewide.

(b) *Communication Check.* The Intoximeters Intox DMT Dual Sensor pushes data to the FTP server via a network connection when a test is completed. If data transmission fails there will be a message displayed on the instrument screen. Each instrument should have at least one file a week uploaded to the FTP site. Open the Data Retrieval and Archiving Program and verify that there is at least one file for every instrument. If there is not at least one file per instrument initiate the necessary corrective actions to restore communication with the instrument missing a file.

(c) *Message Check.* Open the Data Retrieval and Archiving Program to identify the automatic Instrument Messages and Operator Messages retrieved by the current download. Review the messages to determine the needs of the instrument, operator, or location (supplies). Document any actions taken.

(d) *Instrument Parameter Check.* Any instrument parameter that is out of range will trigger a status message. Open the Data Retrieval and Archiving



Program perform a search for status messages. Investigate the source of any problem and document any corrective action.

(e) *Dry Gas Configuration Check.* Open the Data Retrieval and Archiving Program. Review the tank change and DUI tests to verify that the 0.02 Manufacturer, 0.02 Gas Lot Number, 0.02 Gas Expiration Date, 0.08 Manufacturer, 0.08 Gas Lot Number, 0.08 Gas Expiration Date, 0.150 Manufacturer, 0.150 Gas Lot Number, 0.150 Gas Expiration Date, 0.02 Target Concentration, 0.02 Relative Tolerance, 0.02 Absolute Tolerance, 0.08 Target Concentration, 0.08 Relative Tolerance, 0.08 Absolute Tolerance, 0.150 Target Concentration, 0.150 Relative Tolerance, and 0.150 Absolute Tolerance is correct. Investigate any discrepancy and document any corrective action.

(f) *Error and Accuracy Check.* Open the Data Retrieval and Archiving Program perform a search for status messages. Also review at the values obtained during the 0.02 g/210L and 0.08 g/210L standard checks. Evaluate these results for trends. Investigate the source of any problem or trend and document any corrective action.

3. *Periodic Self-Test.* The Periodic Self-Test serves as a way to "exercise" the mechanical components of the Intoximeters Intox DMT Dual Sensor. The Intoximeters Intox DMT Dual Sensor will perform a Periodic Self-Test at designated intervals.

(g) *Filter Wheel Test.* The Filter Wheel Test is part of the Periodic Self-Test that activates the filter wheel and verifies that it is working properly. Any failure will be documented in the Periodic Self-Test file. Review this file for any discrepancy and document any corrective action.

(h) *Five Way Test.* The Five Way Test is part of the Periodic Self-Test that activates the five way valve.

(i) *Bleed Valve Test.* The Bleed Valve Test is part of the Periodic Self-Test that activates the bleed valve on the fuel cell.

4. *Periodic Check.* Serves as a mechanism to exercise the fuel cell of the Intoximeters Intox DMT Dual Sensor at locations where the instrument is rarely used. The Intoximeters Intox DMT Dual Sensor will perform the Periodic Check of the 0.080 g/210L

standard at designated intervals. The results of the Periodic Test will be stored in the Periodic Test file. Review this file for any discrepancy and document any corrective action.

5. Annual Re-Evaluation/Pre-Installation Evaluation. Regardless of the performance of an instrument in the field each instrument will be brought back to the Alabama Department of Forensic Sciences laboratory once a year for evaluation. This evaluation will consist of the battery of tests described under (1) Pre-Installation Evaluation.

**Author:** Dale A. Carpenter, Mark A. Pevey, Gregory L. Turner

**Statutory Authority:** Code of Ala. 1975, §32-5A-194, as amended.

**History:** New Appendix: Filed July 7, 2003; effective August 11, 2003. **Amended:** Filed May 10, 2018; effective June 24, 2018.

**Amended:** Filed May 6, 2019, effective June 20, 2019.

**Ed. Note:** Appendix is attached to Rule 370-1-1-.01, as per certification filed July 7, 2003; effective August 11, 2003.